d	Technical Inspection of Scottish Water Sampling and Analytical Services
	Edinburgh Laboratory 11-13 August 2015
	11-15 August 2015
DWQR Staff Present Matt Bower, Moira Malcolm Scottish Water Staff Present (& Titles) Sharon Reilly and Colin Bain (Quality Team)	
Summary of Inspection	
Overall Summary	
	Excellent skills and knowledge demonstrated throughout the audit by staff at all levels. Within the laboratory, housekeeping was very good and appropriate equipment checks were being undertaken. Good record keeping demonstrated through all processes and chain of custody on transportation of samples was excellent and a further improvement on the already good standard encountered during previous audits.
	Number of Findings: 0
Sampl	ing Frequency Compliance
	Regulatory samples taken at appropriate frequencies and evenly spaced throughout the sampling period. No deficiencies noted.
Sampl	ing
	Sample record keeping at a good standard throughout. Sample techniques not examined.
Trans	portation
	Transportation sufficient, record keeping relating to transport all completed well with excellent chain of custody demonstrated throughout.
Reage	ints and Equipment
	Equipment and reagents all maintained well with excellent record keeping. Maintenance and calibrations being performed satisfactorily. Some temperature control equipment 'sensitive' - see finding 2 below.
Analysis	
	Good analysis techniques demonstrated in all departments. Slight noted issue with a minority of crypto slides - see finding 1 below.
Method and Result Recording	
	Recording practices of a good quality throughout. Moves being made to increase opportunities for direct data capture.
Quality Control	
	Robust quality management systems in all areas. Quality team playing a highly effective role in the maintenance of excellent standards in the laboratory. AQC methods being used effectively across all departments.

Staffing and Training

All staff showed good knowledge of methods and techniques and were enthusiastic throughout. This demonstrates good training and management culture. Staff were being given good access to devlopment opportunities. Laboratory fully compliant with *Competent Analyst* requirements.

Action Following Exceedences

Procedure and practices demonstrated for reporting and progressing exceedences all appropriate and timeous. Good record keeping demostrated throughout.